



ISTITUTO NAZIONALE DI RICERCA METROLOGICA Repository Istituzionale

A correlation noise spectrometer for flicker noise measurement in graphene samples

Original

A correlation noise spectrometer for flicker noise measurement in graphene samples / Marzano, Martina; Cultrera, Alessandro; Ortolano, Massimo; Callegaro, Luca. - In: MEASUREMENT SCIENCE & TECHNOLOGY. - ISSN 0957-0233. - 30:3(2019), p. 035102. [10.1088/1361-6501/aafcab]

Availability:

This version is available at: 11696/61085 since: 2020-01-29T10:09:48Z

Publisher:

Published

DOI:10.1088/1361-6501/aafcab

Terms of use:

Visibile a tutti

This article is made available under terms and conditions as specified in the corresponding bibliographic description in the repository

Publisher copyright

(Article begins on next page)

PAPER • OPEN ACCESS

A correlation noise spectrometer for flicker noise measurement in graphene samples

To cite this article: Martina Marzano *et al* 2019 *Meas. Sci. Technol.* **30** 035102

View the [article online](#) for updates and enhancements.

Recent citations

- [Eliminating friction between flat specimens and an antibuckling support during cyclic tests using a simple sensor](#)
Domen Šeruga *et al*

A correlation noise spectrometer for flicker noise measurement in graphene samples

Martina Marzano^{1,2}, Alessandro Cultrera², Massimo Ortolano^{1,2}
and Luca Callegaro²

¹ Dipartimento di Elettronica e Telecomunicazioni, Politecnico di Torino, Corso Duca degli Abruzzi 24, 10129 Torino, Italy

² INRIM—Istituto Nazionale di Ricerca Metrologica, Strada delle Cacce 91, 10135 Torino, Italy

E-mail: martina.marzano@polito.it

Received 15 October 2018, revised 10 December 2018

Accepted for publication 7 January 2019

Published 7 February 2019



Abstract

We present a high-resolution digital correlation spectrum analyzer for the measurement of low frequency resistance fluctuations in graphene samples. The system exploits the cross-correlation method to reject the amplifiers' noise. The graphene sample is excited with a low-noise DC current. The output voltage is fed to two two-stage low-noise amplifiers connected in parallel; the DC signal component is filtered by a high-pass filter with a cutoff frequency of 34 mHz. The amplified signals are digitized by a two-channel synchronous ADC board; the cross-periodogram, which rejects uncorrelated amplifiers' noise components, is computed in real time. As a practical example, we measured the noise cross-spectrum of graphene samples in the frequency range from 0.153 Hz to 10 kHz, both in two- and four-wire configurations, and for different bias currents. We report here the measurement setup, the data analysis and the error sources.

Keywords: instrumental noise, instrument optimisation, flicker noise, digital signal processing, data acquisition, graphene

(Some figures may appear in colour only in the online journal)

1. Introduction

Like other electrically conducting materials, when crossed by a current, graphene exhibits electrical *excess noise*, dominated by flicker ($1/f$) noise. Flicker noise limits the resolution of sensors [1–4] and the sensitivity of amplifiers and detectors [5–7]. Moreover, in recent years, several works proposed graphene sensors based on electrical noise output [7–10].

Flicker noise in graphene is highly dependent on the growth technique, the device fabrication technology and the specific bias conditions [5, 11–17]. A proper characterisation of magnitude and spectral properties of flicker noise is essential to enable its adoption as an industrial material for future electronics.

The accurate measurement of flicker noise spectral density is made difficult by the small magnitude of the signal to be measured (in the $\text{nV Hz}^{-1/2}$ range) [18], and the long measurement time required to probe the low frequency region [19].

A direct measurement using a single-channel signal analyzer is corrupted by the flicker noise of the instrument itself [20–22], which can have a magnitude comparable to that of the signal of interest. A correction can be performed, but requires an independent measurement of the noise floor [23, 24]. Interferences, often from mains, are another typical source of error.

The experimental design must take into consideration the DC voltage caused by the device bias current, which can overload a DC-coupled instrument input stage. AC coupling in commercial amplifiers is available, but typically with a cutoff frequency of tenths of hertz (see e.g. [25]), causing significant errors on measurements performed below 10 Hz.

This paper presents a digital correlation spectrum analyzer for the measurement of flicker noise of graphene samples.

Original content from this work may be used under the terms of the [Creative Commons Attribution 3.0 licence](#). Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.



The analyzer is based on cross-correlation [26, 27], which rejects to a large extent the noise of the amplifiers, and thus allows to determine the device noise power spectrum under DC current excitation. Both analogue [28, 29] and digital [30, 31] cross-correlators have been described in the literature, the most accurate implementations being employed in Johnson noise thermometry experiments [32–34]. Although not perfect [35–37], the rejection of amplifiers' noise given by correlation is particularly effective at low frequency, and therefore in flicker noise measurements.

The analyzer here described is based on a two-channel voltage signal conditioning system, including low-noise, high-gain amplifiers and a synchronous sampling system. The acquired samples are processed by a digital correlation algorithm. At variance with commercial signal analyzers (see e.g. [25] and [38]), the lowest measurement frequency can be virtually arbitrarily extended, with a number of frequency points up to 2^{17} . Proper wiring and shielding reduce interferences to negligible levels.

In section 4, examples of flicker noise measurements on graphene samples are shown.

2. Noise measurement concepts

2.1. Basics of cross-correlation

Noise measurements on devices require the amplification of small signals. Amplifiers, however, introduce additional noise components. The cross-correlation method rejects the amplifiers' noise by simultaneously amplifying the device noise with two different amplifiers and by combining their output signals in a suitable way.

Let us briefly review the cross-correlation method by referring to the principle schematic of figure 1: $v(t)$ represents the device noise signal, which is the quantity of interest; $e_1(t)$ and $e_2(t)$ are the amplifiers' noise components; the amplifier gains are assumed to be 1, without loss of generality. We assume that the signals are realizations of stationary and ergodic random processes. For this class of processes, the main statistical properties are described by the auto- and cross-correlation functions or, equivalently, in the frequency domain, by the spectral density and cross-spectral density functions (or, respectively, spectrum and cross-spectrum).

The signals $v_1(t)$ and $v_2(t)$ at the amplifiers' outputs are

$$v_1(t) = v(t) + e_1(t), \quad (1)$$

$$v_2(t) = v(t) + e_2(t). \quad (2)$$

We assume that $v(t)$, $e_1(t)$ and $e_2(t)$ are uncorrelated, that is, for all time lags τ ,

$$E\{e_1(t)e_2(t+\tau)\} = 0, \quad (3)$$

$$E\{v(t)e_i(t+\tau)\} = 0, \quad i = 1, 2, \quad (4)$$

where $E\{\cdot\}$ denotes the expected value of the argument. With these assumptions, the cross-correlation of $v_1(t)$ and $v_2(t)$ is

$$\begin{aligned} R_{12}(\tau) &= E\{v_1(t)v_2(t+\tau)\}, \\ &= E\{[v(t) + e_1(t)][v(t+\tau) + e_2(t+\tau)]\}, \\ &= E\{v(t)v(t+\tau)\} + 0, \\ &= R_{vv}(\tau), \end{aligned} \quad (5)$$

which coincides with the auto-correlation function $R_{vv}(\tau)$ of the device noise. The terms depending on the uncorrelated noise components are thus rejected: there only remains the term depending on the correlated noise at the amplifiers' inputs.

Equivalently, in the frequency domain, the cross-spectrum, which is defined as the Fourier transform of the cross-correlation $R_{12}(\tau)$, coincides with the spectrum $S_v(f)$ of the only correlated component. In fact, from (5),

$$\begin{aligned} S_{12}(f) &= \int_{-\infty}^{\infty} R_{12}(\tau) e^{-j2\pi f\tau} d\tau, \\ &= \int_{-\infty}^{\infty} R_{vv}(\tau) e^{-j2\pi f\tau} d\tau = S_v(f). \end{aligned} \quad (6)$$

Typically, to estimate $S_{12}(f)$ and $S_v(f)$, the signals $v_1(t)$ and $v_2(t)$ are periodically and simultaneously sampled and acquired with sampling period T_s . We denote the N acquired samples by $v_1[n]$ and $v_2[n]$, $n = 1, \dots, N$.

We choose the cross-periodogram [39, section 9.5] as an estimator of the cross-spectrum and to reduce the uncertainty we adopt the Bartlett lag window [40, section 6.2]. Let us consider NM samples splitted into M groups of N samples each; the cross-periodogram associated to group m , $m = 1, \dots, M$, is

$$\begin{aligned} \hat{S}_{12,m}^{(p)}(f_k) &= \frac{T_s}{N} \left\{ \sum_{n=1}^N v_1[(m-1)N + n] e^{-j2\pi f_k n T_s} \right\}^* \\ &\times \left\{ \sum_{n'=1}^N v_2[(m-1)N + n'] e^{-j2\pi f_k n' T_s} \right\}, \end{aligned} \quad (7)$$

where $f_k = k/(NT_s)$, $k = 0, \dots, N-1$, are the Fourier frequencies and the asterisk denotes complex conjugation. The reciprocal of the acquisition time $T_0 = NT_s$ corresponds to the resolution bandwidth. The average value of the M periodograms,

$$\hat{S}_{12}(f_k) = \frac{1}{M} \sum_{m=1}^M \hat{S}_{12,m}^{(p)}(f_k), \quad (8)$$

is an estimator of $S_{12}(f)$ whose variance [39, section 9.5]

$$\text{var}\{\hat{S}_{12}(f_k)\} \sim \frac{1}{M} S_{v_1}(f_k) S_{v_2}(f_k) \quad (9)$$

depends on M and on the power spectral densities of the signals $v_1(t)$ and $v_2(t)$,

$$S_{v_j}(f_k) = S_v(f_k) + S_{e_j}(f_k) \quad j = 1, 2. \quad (10)$$

The uncertainty of this estimator thus depends on the amplifiers' noise, and choosing a sufficiently large M allows to reduce the uncertainty to the desired level.




Figure 1. Equivalent circuit of a spectrum analyzer based on the cross-correlation method.




Figure 2. Equivalent circuit for the analysis of systematic errors.

2.2. Systematic error sources

A detailed analysis of the error sources in the cross-correlation method can be found in [36, 37, 41]. Here we briefly summarize the results in a form suitable to estimate the systematic error of the setup described in section 3.

A simplified equivalent circuit for the error analysis is shown in figure 2. The device is represented by a Thévenin's equivalent circuit composed of the noise signal source $v(t)$ in series with the resistance R . The two amplifiers have equal gain, which we assume to be 1, and equal input impedance $Z_i(f)$. The amplifiers' output voltages are $v_1(t)$ and $v_2(t)$. The voltage sources $e_1(t)$ and $e_2(t)$ represent the equivalent input noise voltage of the amplifiers; the current sources $j_1(t)$ and $j_2(t)$ represent the input short-circuit noise current of the amplifiers³.

The currents $j_1(t)$ and $j_2(t)$, crossing R and the impedances Z_i , generate a voltage which adds to the signal of

interest, thus causing a systematic error in the estimation of $S_v(f)$. Following [37], it can be shown that the systematic error $\Delta S_v(f)$ on $S_v(f)$ is given by

$$\Delta S_v(f) = R^2 [S_{j_1}(f) + S_{j_2}(f)] - R \frac{\text{Re}[A(f)S_{e_1j_1}^*(f) + A(f)S_{e_2j_2}^*(f)]}{|A(f)|^2}, \quad (11)$$

where $S_{j_1}(f)$ and $S_{j_2}(f)$ are, respectively, the spectral density functions of $j_1(t)$ and $j_2(t)$, $S_{e_1j_1}(f)$ and $S_{e_2j_2}(f)$ are, respectively, the cross-spectral density functions between $e_1(t)$ and $j_1(t)$ and between $e_2(t)$ and $j_2(t)$,

$$A(f) = \frac{1}{1 + 2R/Z_i(f)} \quad (12)$$

is the input attenuation, and the operator Re takes the real part of the argument.

At low frequency, capacitive effects are negligible and the input impedance is usually very high. As a consequence, $|A(f)| \approx 1$ and the contribution of the voltage noise to $S_{j_1}(f)$ and $S_{j_2}(f)$ is usually negligible (see also [37]). The

³ The short-circuit noise current contains also the current injected by the noise voltage through the finite input impedance: with this choice, the input impedance Z_i can be moved upstream of the noise voltage sources, and the equation can be simplified.




Figure 3. Block schematic of the digital correlation spectrum analyzer, composed of a test circuit, amplifiers and data acquisition board with analogue-to-digital converters. The test circuit and the amplifiers are individually electrically shielded (thin rectangles) and further magnetically shielded by a μ metal box (thick rectangle).




Figure 4. Schematic of the test circuit.

cross-spectra $S_{e_{1j_1}}(f)$ and $S_{e_{2j_2}}(f)$ are of difficult evaluation, but at low frequency their contribution is usually negligible too. Taking into account these conditions, we shall approximate the error at low frequency as

$$\Delta S_v(f) \approx R^2[S_{j_1}(f) + S_{j_2}(f)]. \quad (13)$$

3. Measurement setup

The voltage noise of a device or a circuit can be measured with a digital correlation spectrum analyzer in a shielded and temperature controlled environment. Here the device is a graphene sample in the form of a Hall bar. If the sample is unbiased, the measured signal includes only the thermal, white noise component. Instead, if the sample is crossed by a DC bias current, both white and excess noise components are measured.

The measurement setup, as shown diagrammatically in figure 3, is composed of a test circuit, amplifiers and a data acquisition board. Here we describe the setup in a two-terminal configuration; the extension to the four-terminal configuration is straightforward.

The test circuit, represented in figure 4, allows to measure the noise signal $v(t)$ of the sample with or without a DC bias current, which can be turned on and off by the switch S. The bias current is generated by the voltage source E in series with the 10 MΩ metal resistor R_B , having negligible excess noise. The voltage source E consists of four series-connected 1.2V NiCd batteries and a dip switch allowing to set E to 1.2V, 2.4V, 3.6V or 4.8V. The two outputs, 1 and 2, are AC coupled by two high-pass filters with a cutoff frequency of 34 mHz. The test circuit is shielded by a box connected to the low terminal of the battery.

The noise signal $v(t)$ at the outputs 1 and 2 is then amplified simultaneously by two two-stage low noise amplifiers with a total gain of 10^4 (figure 3). The first stage is composed of two battery-powered commercial amplifiers with a gain of 100: an EG&G PAR 113 pre-amplifier (10 nV Hz^{-1/2} nominal voltage noise, 7 fA Hz^{-1/2} current noise at 1 kHz) and an EG&G PAR 5113 pre-amplifier (4 nV Hz^{-1/2} nominal voltage noise, 40 fA Hz^{-1/2} current noise at 1 kHz). The second stage is composed of two bespoke amplifiers with a gain of 100 (10 nV Hz^{-1/2} voltage noise), powered by a dedicated regulated power supply.




Figure 5. Measurement setup: (a) test circuit placed in an electrically shielding box (top cover removed); (b) test circuit and two-stage amplifiers placed in a large magnetically shielding μmetal box (top cover removed).




Figure 6. (a) Scanning electron microscope image of the graphene Hall bar AEM23C; (b) two-terminal measurement configuration: the noise voltage $v(t)$ is measured across the same terminals at which the DC bias current I is applied; (c) four-terminal measurement configuration: the noise voltage is measured across two terminals which are different from those at which the DC bias current is applied. The dashed lines represent the bias current path in the respective cases.

The test circuit (figure 5(a)) and the two two-stage amplifiers are contained in an electrically and magnetically shielded μmetal box, as shown in figure 5(b). The amplifiers' outputs are connected to a PXI rack, equipped with the data acquisition board (National Instruments 4462) that communicates through an optical fiber with a computer. The board is characterized by a resolution of 24 bit and a maximum sampling frequency of 204.8 kHz. A software controls the signal acquisition. A typical measurement is composed of $N = 2^{17}$ pairs of samples acquired at a sampling rate of 20 kHz. The resolution bandwidth is 0.153 Hz and the duration T_0 of a single acquisition is of about 6.5 s. The shielded boxes and the amplifiers have the same ground connected to the PXI external ground.

4. Example measurements on graphene samples

As an example application of the setup described in section 3, we present a series of noise measurements on two graphene samples, labeled respectively AEM23C and AEM22A (see [17] for an in-depth analysis of the measurements). The Hall bars were fabricated from commercial chemical vapor deposited monolayer graphene. The geometry was defined by means of electron beam lithography and oxygen plasma etching. Cr/Au electrodes were deposited with an electron-gun evaporator. Figure 6(a) shows a scanning electron microscope image of the sample AEM23C.

Noise measurements were performed in two- and four-terminal configurations (see figures 6(b) and (c)), and at




Figure 7. Voltage noise spectra of sample AEM23C in different measurement configurations: (a) estimated spectra $\hat{S}_{v_1}(f)$ and $\hat{S}_{v_2}(f)$ measured in two-terminal configuration without bias and the corresponding cross-spectrum $\hat{S}_{12}(f)$; (b) comparison between two-terminal (2T) and four-terminal (4T) configurations with (on) or without (off) DC bias current.

different bias current levels. The two-terminal configuration corresponds to the schematic of figure 4.

Figure 7(a) shows the successful rejection of uncorrelated noise by the cross-correlation technique. In this figure, three voltage noise spectra from sample AEM23C are reported. The two spectra labelled $\hat{S}_{v_1}(f)$ and $\hat{S}_{v_2}(f)$ were measured in the two-terminal configuration with no bias current and represent the measurements at the two amplifiers' outputs (scaled by the gain). The spectrum labelled $\hat{S}_{12}(f)$ is the estimated cross-spectrum of $\hat{S}_{v_1}(f)$ and $\hat{S}_{v_2}(f)$. All the estimated spectra were obtained by averaging $M = 250$ periodograms. The spectra $\hat{S}_{v_1}(f)$ and $\hat{S}_{v_2}(f)$ contain the thermal noise from the unbiased sample but also the amplifiers' noise, with $1/f$ components. These components are uncorrelated and are rejected in the cross-spectrum $\hat{S}_{12}(f)$ as expected from (5).

Figure 7(b) reports four cross-spectra, obtained in both two-terminal (2T) and four-terminal (4T) configurations and with (on) or without (off) bias current. The cross-spectra from the unbiased sample consist of the thermal noise component only; the thermal noise level measured in two- and four-terminal configurations is the same because the equivalent resistance of the sample as seen from the voltage terminals is the same in both configurations (figure 6). Instead, the cross-spectra from the biased sample contain also a $1/f$ noise component, and the $1/f$ noise level in the two-terminal configuration is one order of magnitude greater than that in the four-terminal configuration. This ratio is due to the fact that in the two-terminal configuration the bias current crosses the c-shaped graphene segment (dashed line in figure 6(b)) and the contacts, whereas in the four-terminal configuration the bias current crosses only the center bar of the sample (dashed line in figure 6(c)) and only a fraction of the generated $1/f$ noise is thus measured. Given the sample geometry and the bias current paths, if the excess noise were generated in the graphene only, there would have been a ratio of

about 3 between the $1/f$ noise levels measured in two- and four-terminal configurations. The extra factor can be ascribed to the contacts [17].

The systematic error due to the amplifiers' input currents can be estimated from (13) and from the amplifiers' current noise specifications reported in section 3. It is worth noting that for FET-input amplifiers, like those chosen for this experiment, the current noises $S_{j_1}(f)$ and $S_{j_2}(f)$ are white down to very low frequency, and this implies that the systematic error is constant also in the flicker noise region of interest. For a resistance of the order of 10 k Ω , $\Delta S_v(f) \approx 1.6 \times 10^{-19} \text{ V}^2 \text{ Hz}^{-1}$, independent of frequency. The type A uncertainty associated to the spectra can be evaluated from (9). For instance, from figure 7(a),

$$\hat{S}_{v_1}(f) \approx 2 \times 10^{-16} \text{ V}^2 \text{ Hz}^{-1} \times \left(1 + \frac{20 \text{ Hz}}{f}\right) \quad (14)$$

and

$$\hat{S}_{v_2}(f) \approx 3 \times 10^{-16} \text{ V}^2 \text{ Hz}^{-1} \times \left(1 + \frac{20 \text{ Hz}}{f}\right), \quad (15)$$

from which, taking into account that $M = 250$, $u(\hat{S}_{12}(f)) \approx 1.5 \times 10^{-17} \text{ V}^2 \text{ Hz}^{-1}$ in the white noise region and $u(\hat{S}_{12}(f)) \approx 3 \times 10^{-16} \text{ V}^2/f$ in the $1/f$ noise region. A similar analysis can be performed for figure 7(b). Indeed, the type A uncertainty can be further reduced by increasing M .

To give a further example, we report in figure 8(a) a number of measurements on the sample AEM22A, in two-terminal configuration at bias current levels of approximately 120 nA, 240 nA, 360 nA, 480 nA. Also in this case the rejection of uncorrelated noise is successful for all the spectra. It can be observed that the flicker noise level increases with the bias current. The expected quadratic dependence [42, 43] is confirmed by figure 8(b), which reports the spectra normalized to the square of the corresponding applied current.




Figure 8. Voltage noise spectra of sample AEM22A at different bias current levels I in two-terminal configuration (a). The same spectra were normalized to the corresponding bias current levels and their low frequency parts are shown in (b).

We finally remark upon the virtual absence of 50 Hz power-line disturbances from the spectra of figures 7(a) and (b), showing the effectiveness of the shielding arrangement employed in the experiment (in figure 8(a), the level of 50 Hz disturbances is slightly higher because, at the time of that measurement, the grounding scheme was not yet optimized).

5. Conclusions

The spectrum analyzer described, based on digital sampling and a cross-correlation algorithm, allows measurements of excess noise at frequencies down to the 100 mHz range, limited only by the cutoff frequency of the high-pass filters employed. The whole sequence of samples is recorded for off-line processing, allowing further statistical analyses to be performed, or the selective removal of outliers generated by burst interferences.

A model of the residual correlation effects, which can give rise to systematic noise errors, has been developed. The model and the experiment confirm that, whereas the employed general-purpose front-end amplifiers have a large flicker noise voltage, the resulting analyzer noise floor is white even at very low frequency.

Excess noise spectra from graphene samples, either in two-terminal or four-terminal configuration and at different bias current levels, were measured. The noise generated by a graphene layer and by the graphene-metal contacts was measured with high resolution, allowing accurate quantitative determinations of the flicker noise exponent, and of its magnitude dependence on the excitation current. These parameters are of special interest for the proper identification of the noise generation mechanisms as shown in [17].

Acknowledgments

The work has been realized within the Joint Research Project 16NRM01 GRACE: Developing electrical characterisation

methods for future graphene electronics. This project has received funding from the EMPIR programme co-financed by the Participating States and from the European Union's Horizon 2020 research and innovation programme.

ORCID iDs

Martina Marzano <https://orcid.org/0000-0001-5288-3093>
 Alessandro Cultrera <https://orcid.org/0000-0001-8965-9116>
 Massimo Ortolano <https://orcid.org/0000-0002-7217-8276>
 Luca Callegaro <https://orcid.org/0000-0001-5997-9960>

References

- [1] Xu H, Huang L, Zhang Z, Chen B, Zhong H and Peng L-M 2013 Flicker noise and magnetic resolution of graphene Hall sensors at low frequency *Appl. Phys. Lett.* **103** 112405
- [2] Grover S, Dubey S, Mathew J P and Deshmukh M M 2015 Limits on the bolometric response of graphene due to flicker noise *Appl. Phys. Lett.* **106** 051113
- [3] Wang T, Huang D, Yang Z, Xu S, He G, Li X, Hu N, Yin G, He D and Zhang L 2016 A review on graphene-based gas/vapor sensors with unique properties and potential applications *Nano-Micro Lett.* **8** 95–119
- [4] El-Ahmar S, Koczorowski W, Poźniak A A, Kuświk P, Strupiński W and Czajka R 2017 Graphene-based magnetoresistance device utilizing strip pattern geometry *Appl. Phys. Lett.* **110** 043503
- [5] Hossain M Z, Romyantsev S, Shur M S and Balandin A A 2013 Reduction of $1/f$ noise in graphene after electron-beam irradiation *Appl. Phys. Lett.* **102** 153512
- [6] Huang L, Zhang Z, Chen B, Ma X, Zhong H and Peng L-M 2014 Ultra-sensitive graphene Hall elements *Appl. Phys. Lett.* **104** 183106
- [7] Amin K R and Bid A 2015 High-performance sensors based on resistance fluctuations of single-layer-graphene transistors *ACS Appl. Mater. Interfaces* **7** 19825–30
- [8] Romyantsev S, Liu G, Shur M S, Potyrailo R A and Balandin A A 2012 Selective gas sensing with a single pristine graphene transistor *Nano Lett.* **12** 2294–8

- [9] Rumyantsev S, Liu G, Potyralo R A, Balandin A A and Shur M S 2013 Selective sensing of individual gases using graphene devices *IEEE Sens. J.* **13** 2818–22
- [10] Samnakay R, Jiang C, Rumyantsev S, Shur M and Balandin A 2015 Selective chemical vapor sensing with few-layer MoS₂ thin-film transistors: comparison with graphene devices *Appl. Phys. Lett.* **106** 023115
- [11] Lu J, Pan J, Yeh S-S, Zhang H, Zheng Y, Chen Q, Wang Z, Zhang B, Lin J-J and Sheng P 2014 Negative correlation between charge carrier density and mobility fluctuations in graphene *Phys. Rev. B* **90** 085434
- [12] Balandin A A 2013 Low-frequency 1/f noise in graphene devices *Nat. Nanotechnol.* **8** 549–55
- [13] Amin K R and Bid A 2015 Effect of ambient on the resistance fluctuations of graphene *Appl. Phys. Lett.* **106** 183105
- [14] Arnold H N, Sangwan V K, Schmucker S W, Cress C D, Luck K A, Friedman A L, Robinson J T, Marks T J and Hersam M C 2016 Reducing flicker noise in chemical vapor deposition graphene field-effect transistors *Appl. Phys. Lett.* **108** 073108
- [15] Rumyantsev S L, Shur M S, Liu G and Balandin A A 2017 Low frequency noise in 2D materials: graphene and MoS₂ *Int. Conf. on Noise and Fluctuations* pp 1–4
- [16] Kochat V, Tiwary C S, Biswas T, Ramalingam G, Hsieh K, Chattopadhyay K, Raghavan S, Jain M and Ghosh A 2016 Magnitude and origin of electrical noise at individual grain boundaries in graphene *Nano Lett.* **16** 562–7
- [17] Cultrera A, Callegaro L, Marzano M, Ortolano M and Amato G 2018 Role of plasma-induced defects in the generation of 1/f noise in graphene *Appl. Phys. Lett.* **112** 093504
- [18] Buckingham M J 1985 *Noise in Electronic Devices and Systems* (New York: Horwood)
- [19] Rubiola E, Francese C and Marchi A D 2001 Long-term behavior of operational amplifiers *IEEE Trans. Instrum. Meas.* **50** 89–94
- [20] Martin S S 1999 Method and apparatus for measuring electrical noise in devices *US Patent* 5970429
- [21] Blaum A, Pilloud O, Scalea G, Victory J and Sischka F 2001 A new robust on-wafer 1/f noise measurement and characterization system *Proc. 2001 Int. Conf. on Microelectronic Test Structures*
- [22] Giusi G, Giordano O, Scandurra G, Ciofi C, Rapisarda M and Calvi S 2015 Automatic measurement system for the DC and low-f noise characterization of FETs at wafer level *Proc. 2015 IEEE Int. Instrumentation and Measurement Technology Conf.* (<https://doi.org/10.1109/I2MTC.2015.7151606>)
- [23] Voss R F and Clarke J 1976 Flicker (1/f) noise: equilibrium temperature and resistance fluctuations *Phys. Rev. B* **13** 556–73
- [24] Macucci M and Pellegrini B 1991 Very sensitive measurement method of electron devices current noise *IEEE Trans. Instrum. Meas.* **40** 7–12
- [25] Stanford Research Systems 2018 SR785—100kHz two-channel dynamic signal analyzer *Data Sheet*
- [26] Rubiola E 2008 The magic of cross correlation in measurements from dc to optics *Proc. 22nd European Frequency and Time Forum (Toulouse, France 22–25 April 2008)*
- [27] White D R et al 1996 The status of Johnson noise thermometry *Metrologia* **33** 325
- [28] Fink H J 1959 A new absolute noise thermometer at low temperatures *Can. J. Phys.* **37** 1397–406
- [29] Brophy J J, Epstein M and Webb S L 1965 Correlator-amplifier for very low level signals *Rev. Sci. Instrum.* **36** 1803–6
- [30] Sampietro M, Fasoli L and Ferrari G 1999 Spectrum analyzer with noise reduction by cross-correlation technique on two channels *Rev. Sci. Instrum.* **70** 2520–5
- [31] Ferrari G and Sampietro M 2002 Correlation spectrum analyzer for direct measurement of device current noise *Rev. Sci. Instrum.* **73** 2717–23
- [32] Callegaro L, D'Elia V, Pisani M and Pollarolo A 2009 A Johnson noise thermometer with traceability to electrical standards *Metrologia* **46** 409
- [33] Flowers-Jacobs N E, Pollarolo A, Weiss A, Coakley K, Fox A E, Rogalla H, Tew W L and Benz S P 2017 The NIST Johnson noise thermometry system for the determination of the Boltzmann constant *J. Res. Natl Inst. Stand. Technol.* **122** 1–43
- [34] Flowers-Jacobs N E, Pollarolo A, Coakley K J, Fox A E, Rogalla H, Tew W L and Benz S P 2017 A Boltzmann constant determination based on Johnson noise thermometry *Metrologia* **54** 730
- [35] White D R 1984 Systematic errors in a high-accuracy Johnson noise thermometer *Metrologia* **20** 1
- [36] White D R and Zimmermann E 2000 Preamplifier limitations on the accuracy of Johnson noise thermometers *Metrologia* **37** 11
- [37] Callegaro L, Pisani M and Ortolano M 2010 Systematic errors in the correlation method for Johnson noise thermometry: residual correlations due to amplifiers *Metrologia* **47** 272
- [38] Keysight 2018 35670A FFT dynamic signal analyzer, DC-102.4kHz *Data Sheet*
- [39] Priestley M B 1981 *Spectral Analysis and Time Series* vol 2 (London: Academic)
- [40] Priestley M B 1981 *Spectral analysis and time series* vol 1 (London: Academic)
- [41] White D R, Benz S P, Labenski J R, Nam S W, Qu J F, Rogalla H and Tew W L 2008 Measurement time and statistics for a noise thermometer with a synthetic-noise reference *Metrologia* **45** 395–405
- [42] Hooge F N 1976 1/f noise *Physica B + C* **83** 14–23
- [43] Vandamme L K J 2013 How useful is Hooge's empirical relation *Int. Conf. on Noise and Fluctuations (IEEE)* (<https://doi.org/10.1109/ICNF.2013.6578875>)